



KEFRI/AD/SD/VOL.XI (355)

14TH NOVEMBER 2023

TO ALL INTERESTED BIDDERS

Dear All,

CLARIFICATION/ADDENDUM NO. 1

TENDER NO. KEFRI/ONT/002/2023-2024

TENDER NAME: SUPPLY, DELIVERY, INSTALLATION AND COMMISSIONING OF ASSORTED LABORATORY AND FIELD EQUIPMENT FOR ELEVEN (11) KEFRI SEED CENTRES

The Kenya Forestry Research Institute (KEFRI) would wish to bring to your attention the below clarification to the above tender advertised on MyGov. – Issue No. 19/2023-2024, which appeared in the Standard Newspaper of **Tuesday, 7th November, 2023.**

1. TENDER DATA SHEET

The below ITT are clarified as follows.

ITT Reference	Particulars of Appendix To Instructions To Tenders
ITT 13.6	Prices quoted shall correspond to 100 % of the items specified for each lot and to 100% of the quantities specified for each item of a lot.
ITT 33.4	<ul style="list-style-type: none">i) If a tenderer wins only one Lot, the tenderer will be awarded a contract for that Lot, provided the tenderer meets the Eligibility and Qualification Criteria for that Lot.ii) If a tenderer wins more than one Lot, the tenderer will be awarded contracts for all won Lots, provided the tenderer meets the aggregate Eligibility and Qualification Criteria for all the Lots.

2. EVALUATION CRITERIA

2.1 Stage 1 – Preliminary Examination

It's clarified that MR 11 shall read as follows:

Mandatory Requirement	As is in the Tender Document	Clarification/Requirement
MR 11	Must attach a Valid Direct Manufacturers Authorization letters/certificate if not the manufacturer or an authorization letter from a distributor who holds a manufacturer authorization. The manufacturer authorization to the distributor to be attached as well.	Bidders are required to attach a Valid Direct Manufacturers Authorization letters/certificate if not the manufacturer or an authorization letter from a distributor who holds a manufacturer authorization for the following equipment: <ol style="list-style-type: none">1. Seed cleaner (Blower)2. Autoclave3. Drying oven (0 -240 oC)4. Germination Cabinets5. High precision microscope6. Lamina Flow7. Seed counter8. Marvin seed analyzer

3. ADDITIONAL REQUIREMENT AND CHANGES TO THE PRICE SCHEDULE.

The following changes are made to the Price Schedule -Seed processing and testing as indicated below.

- i. Lot 13 is changed from Seed cleaner to Seed counter.
- ii. Lot 15 is introduced.


PRICE SCHEDULE

LOTS		Description	Country of Origin/Assembled /Manufacturer	UoM	Qty	Location to be delivered	Unit Cost	Total Price per Line-item Incl. of VAT
		SEED PROCESSING AND TESTING.						
Lot 13	1	Seed Counter		No	6	Headquarter		
Lot 15	1	Seed huller		No	13	Headquarter		
		TOTAL						

4. SECTION IV -SCHEDULE OF REQUIREMENTS

Item No. 17 on the scheduled of requirements is changed from Seed Cleaner to Seed Blower as below.

DETAILED TECHNICAL SPECIFICATIONS

No	Equipment	Equipment Specification	Compliance to KEFRI Specifications -YES/NO
17		<ul style="list-style-type: none"> i. Column cylinder seed blower ii. Quick and easy separation of light and heavy seed iii. Dimensions: H x W x D: 1045 mm x 440 mm x 220 mm iv. Weight: 7.5 kg v. At least 3 collecting vessels installed on the column cylinder. vi. Power: At least 0.33HP motor vii. Speed: 2850 RPM viii. Voltage: 220-240V; 50/60Hz 	

NOTE – This Clarification/Addendum -No. 1 forms part of the Tender requirements and is binding on all bidders.

All other details, terms and conditions **including closing date and time** remain the same.

For:

HEAD OF SUPPLY CHAIN MANAGEMENT - KEFRI

