

KENYA FORESTRY RESEARCH INSTITUTE

Tel: +254 20 2010651/2
+254 722 157 414
+254 724 259 781/2
+254 734 251 888
Email: director@kefri.org
Website: www.kefri.org

P. O. Box 20412
00200, Nairobi
KENYA



Ref: **KEFRI/AD/SD/VOL.X/(619)**

Date: **25TH APRIL, 2019**

TO: ALL BIDDERS

RE: ADDENDUM NO.1: TENDER NO. KEFRI/HQT/037/2018-2019 – SUPPLY & DELIVERY OF LABORATORY EQUIPMENT

Following the advertisement of the above tender, the Kenya Forestry Research Institute has received requests for clarifications on the tender. Consequently, the Institute would like to provide clarification as follows:

S/No.	Issue	Clarification
1.	Analytical Balance- You have indicated that the balance should be supplied with a rechargeable NiMH Battery Pack. Please note that the analytical balance does not use batteries as it is not portable and not supposed to be moved. Moving the balance will compromise the calibration	The AC Adaptor 240V 50/60Hz is the primary. The battery pack is a power back-up intended to be used in cases where there is no power source.
2.	Top Pan Balance – Kindly clarify on the capacity of the Balance as you have given two capacities, 600gms and 2600 gms	The Top Pan Balance required is 2600gms, not 600gms.
3.	Are bidders supposed to quote for all the items in each lot or can they choose what they can offer from each lot?	Bidders are required to quote for all items in a lot.
4.	In the technical evaluation criteria, you have asked bidders to provide a Certificate of Audit from the Auditor General for the past two years. Kindly clarify and/or confirm that you require Audited Accounts for the past two years, 2016 & 2017	Bidders can provide Certificate of Audit or Audited Accounts for 2016 & 2017
5.	Separation of items	Bidders should tender for the items as indicated in each lot. There shall be no separation.

The other details of the tender remain the same.

H. W. Barasa
Deputy Director, Supply Chain Management

All communications should be addressed to the Director