



KEFRI/AD/SD/VOL XII/95

2/12/2022

TO POTENTIAL BIDDERS

Dear All,





Clarification/Addendum No. 1

Tender No. KEFRI/ONT/008/2022-2023

Tender Name. SUPPLY, DELIVERY, INSTALLATION AND COMMISSIONING OF LABORATORY EQUIPMENT

The Kenya Forestry Research Institute (KEFRI) would wish to bring to the attention of all potential bidders the following clarification;

S/No.	Bidder's Question	KEFRI's Response
1.	We are honored to have been invited to participate in this tender. We did notice some missing specifications, particularly the i) Autoclaves capacity ii) The total number of holes required for each tray in the germination cabinets iii) Preferred tray color of the same.	i) The autoclave load capacity is specified at 20 litres. ii) The germination cabinets should have withdrawable and adjustable racks iii) No preferred color
2.	There are disparities in the technical specifications for the following items: Drying Oven (0-240 OC) and the High Precision Electronic Analytical Balance from the sections: PRICE SCHEDULE AND Detailed Technical Specifications and Standards in SECTION V of the tender document. Please clarify which specifications should be adopted.	The Detailed Technical Specifications and Standards in SECTION V to be adopted.
3.	Drying oven. Please clarify the type of ventilation whether it is forced air circulation(Fan) or natural air circulation(Convection)	Natural air circulation (Convection)

4.	<p>Germination Tray High-density plastic. Kindly share with us the image to enable us to quote accurately.</p> <p>Germination Boxes.</p>	<p>Germination Boxes 1200ml</p> 
		<p>Germination Boxes 600ml</p> 
		<p>Germination Boxes 6300ml</p> 
		<p>Germination Tray</p> 

NOTE – This Clarification/ Addendum (No. 1) forms part of the tender requirements and is binding to all bidders.

All other details, terms and conditions (including closing date and time) remain the same.



Isaac Odhiambo
FOR: Manager –Supply Chain Management